



SETTING THE FORCE

- 1** Turn on the Paraglide remote. Hold the Reset and Data button simultaneously for 3 seconds to open the settings menu.
- 2** Press the Data button to select "Record Force".
- 3** Immediately reposition the user by pressing and holding the Glide button until the user is properly seated in their wheelchair.
- 4** The proper force is now set.

